

¹Olomu P, ¹Hoang S, ²Szmuk P, ¹Skinner M

¹Utsw, Dallas , Dallas , TX, USA; ²Utsw , Dallas , TX, USA

A 10 year old, 36 Kg, female with a locally metastatic papillary thyroid carcinoma presented for total thyroidectomy and central neck dissection. Neck ultrasound showed a nodule measuring 1.9 x 1.5 x 1.2 cm in the left thyroid lobe. Following premedication, anesthesia was induced with sevoflurane and an IV catheter was placed. The trachea was intubated with a 6.0 mm cuffed endotracheal tube with a LSE 500 ms Dragonfly single channel electrode (Neurovision Medical Products, Ventura, CA, USA). Correct electrode position was confirmed using videolaryngoscopy. The electrode was connected to the Nerveäna monitoring unit. The output from the monitoring unit was visualized on a laptop with an EMG Viewer™ software program. Anesthesia was maintained with propofol and dexmedetomidine infusions and fentanyl. In addition to the standard monitoring for nerve detection, the Nerveäna Power Index (NPI) was utilized to trend recurrent laryngeal nerve function. The NPI values at baseline and at thirty minutes after baseline are shown in Table 1. Upon surgery completion, both RLNs were stimulated proximally and distally relative to the glottis and final NPI values are shown in Table 1. Extubation was uneventful and vocal cords movement was normal.

Discussion

The Nerveäna™ System has a new tool for prevention of nerve injury during surgery called the “Nerveäna Power Index.” If all other factors are the same, the repeated stimulation of the RLN, especially at the same anatomic spot, can indicate by loss of power of the muscle response that Neuropraxia is occurring. The Nerveäna System integrates the evoked response, determining the presence of a Compound Motor Action Potential response by an empirically derived CAP Threshold. The “power” of a nerve/muscle response is best determined by integration of the displayed evoked EMG (the area under the waveform). By displaying the CAP integral as a percentage of the CAP threshold, the Nerveäna system creates an Index (Nerveäna Power Index, NPI™) of the power of nerve/muscle response as a percentage of the threshold level of response. The equation for the NPI is as follows: Neuropraxia Power Index (%) = 100 x Event CAP Integration Value / Integration Threshold Value. Repeated stimulation of the nerve under the same conditions provides the surgeon with a real-time index of Neuropraxia (Table 2). The NPI, being quantitative, is much more valuable than simple loss of stimulated responses. In practice, the healthy RLN when first stimulated delivers approximately 450%. In the presence of nerve fatigue the number drops throughout the case. A response approaching 100% should be investigated for possible impending nerve injury.

Initial NPI Values	R VC	L VC
Baseline	366	458
Baseline + 30 mins	327	421
Final NPI values	Right RLN	Left RLN
Proximal	389	431
Distal	378	333

